PrymaLab
PRYMALAB ANALYTICAL · IN-HOUSE LABORATORY
3705 Music Branch Rd, Ashland, KY 41102 · +1 (606) 465-9063 · prymalab.net

CERTIFICATE NO.
PL-26-00155
VERIFICATION KEY
DVM5LDJS
Scan or enter both at prymalab.net/testing/verify

CERTIFICATE OF ANALYSIS · LOT RELEASE

Cagrilintide+Semaglutide 2.5+2.5 mg

Lot **PL-2610-033** · Lyophilized co-blend powder · Label claim **2.5 + 2.5 mg/vial**

TESTED IN-HOUSE BY PRYMALAB
In-house analytical result, not an independent third-party test

LOT / BATCH NUMBER	DATE MADE	DATE OF ANALYSIS	DATE ISSUED
PL-2610-033	2026-10-01	2026-10-02	2026-10-16
RETEST / EXPIRY	**UNITS IN LOT / SAMPLED**	**STORAGE**	**CERTIFICATE TYPE**
2028-10-01	281 / 2	−20 °C, desiccated, dark	Lot COA, 2 pages

COMPOUND
Cagrilintide / Semaglutide Blend
Cagrilintide + Semaglutide co-lyophilized 1:1 blend

CAS NUMBER
1415456-99-3 · 910463-68-2

MOLECULAR FORMULA
—

THEORETICAL MASS (AVG.)
4,409.00 · 4,113.58 Da (Cagrilintide / Semaglutide)

COMPOSITION AND SEQUENCES
Cagrilintide: acylated amylin analog Cys3-Cys8 · **Semaglutide:** acylated GLP-1 analog

PURITY · RP-HPLC, AREA %
99.18%
Lowest component (Semaglutide) • specification ≥ 99.0 % for every component


IDENTITY · COMPONENTS CONFIRMED
2 / 2
Components confirmed by LC-MS within ± 1.0 Da: 4,409.06 / 4,113.65 Da

OVERALL CONCLUSION
PASS
Meets every specification for the tests performed

TEST	METHOD	SPECIFICATION	RESULT	CONCLUSION
Cagrilintide: purity · largest impurity · mass	RP-HPLC 214 nm (area %); LC-MS, ESI+	≥ 99.0 % · ≤ 1.0% · ± 1.0 Da	99.35 % · 0.29 % · 4,409.06 Da	**PASS**
Semaglutide: purity · largest impurity · mass	RP-HPLC 214 nm (area %); LC-MS, ESI+	≥ 99.0 % · ≤ 1.0% · ± 1.0 Da	99.18 % · 0.22 % · 4,113.65 Da	**PASS**
Water content	Karl Fischer	≤ 5.0 %	1.57%	**PASS**
Bacterial endotoxin	LAL, kinetic	≤ 5.0 EU/mg	0.376 EU/mg	**PASS**
Sterility	USP <71>, 14-day	No growth	No growth	**PASS**

1 HPLC CHROMATOGRAM

Data file: PL-2610-033-HPLC.raw · Column C18, 150 × 4.6 mm · UV 214 nm



Peak plot drawn from the integrated peak table (area %, log scale); it is not the instrument trace, which is retained in PL-2610-033-HPLC.raw. † Nominal reference retention time for the assigned impurity; the main-peak RT is observed in this lot. Area % is per component (each component main peak + its impurities = 100%).

For laboratory research use only. Not for human or veterinary use.

2 HPLC METHOD · PL-2610-033-HPLC.RAW

INSTRUMENT
Waters Alliance e2695

DATA SYSTEM
Empower 3

DETECTOR
UV absorbance

DETECTION WAVELENGTH
214 nm

COLUMN
C18, 150 × 4.6 mm, 3.5 µm

MOBILE PHASE A
0.1 % TFA in water

MOBILE PHASE B
0.1 % TFA in MeCN

GRADIENT
5–95 % B over 25 min (linear)

FLOW RATE
1.0 mL/min

COLUMN TEMPERATURE
40 °C

INJECTION VOLUME
10 µL

3 GRADIENT PROGRAM

TIME (MIN)	% A	% B
0.0	95	5
25.0	5	95

4 PEAK TABLE

PEAK	RT (MIN)	AREA	AREA %
1	5.57 †	—	0.20
2	6.96 †	—	0.18
3	7.27 †	—	0.22
4	7.57 †	—	0.22
5	**8.04**	**—**	**99.18**
6	10.50 †	—	0.19
7	11.00 †	—	0.29
8	11.59	—	99.35
9	12.40 †	—	0.17
Total		—	100.00 per component

Main peaks: 5 Semaglutide · 8 Cagrilintide; all other peaks are the assigned impurities of the neighbouring component · area % within each component (main + impurities = 100% per component) · † nominal reference RT · Purity = main peak area ÷ total integrated area × 100, to two decimals; main peak shaded.

5 IDENTITY BY MASS SPECTROMETRY

INSTRUMENT
Waters SQD

IONIZATION / MODE
ESI, positive

DECONVOLUTED MASS
4,409.06 / 4,113.65 Da

THEORETICAL MASS
4,409.00 / 4,113.58 Da

ORDER
Cagrilintide / Semaglutide

DIFFERENCE
+0.06 / +0.07 Da

DATA FILE
PL-2610-033-MS.d

ACCEPTANCE
Within ± 1.0 Da (each component)

6 ADDITIONAL TESTS

TEST	PERFORMED BY	DATE	REPORT / FILE
Water content (Karl Fischer)	PrymaLab Analytical	2026-10-02	—
Endotoxin (LAL, kinetic)	PrymaLab Analytical	2026-10-02	—
Sterility (USP <71>, day-14 read-out)	PrymaLab Analytical	2026-10-16	—

A test with no row here was not performed on this lot. Outside-laboratory reports, when used, are attached and named below.

7 REVIEW AND APPROVAL

D. Reynolds
Analyst · performed the testing
Date: 2026-10-16

S. Calloway
Reviewer · checked against raw data
Date: 2026-10-16

M. Phelps
Lab Director · approved for release
Date: 2026-10-16

Raw data retained at: Empower 3, project PL-2610, archive PL-ARC-2026; 7 years. Outside laboratory reports attached: None

This certificate applies only to the lot named above and describes the units sampled from it; it is not a guarantee for any other lot. Results were obtained by PrymaLab LLC's own laboratory (PrymaLab Analytical) and are not an independent third-party test unless an outside report is attached. Reference values (CAS, formula, theoretical mass, sequence) are literature values for the named compound. Authenticity: scan the QR code or enter certificate number PL-26-00155 and key DVM5LDJS at prymalab.net/testing/verify; any copy that does not match the record shown there is not a PrymaLab certificate. PrymaLab LLC products are supplied for laboratory research use only and not for human or veterinary use.