**PrymaLab**
PRYMALAB ANALYTICAL · IN-HOUSE LABORATORY
3705 Music Branch Rd, Ashland, KY 41102 · +1 (606) 465-9063 · prymalab.net

CERTIFICATE NO.
**PL-26-00323**
VERIFICATION KEY
**222GLLFM**
Scan or enter both at prymalab.net/testing/verify

CERTIFICATE OF ANALYSIS · LOT RELEASE

# Thymogen 10 mg Pen

Lot **PL-2610-211** · Solution · Label claim **10 mg/mL**

**TESTED IN-HOUSE BY PRYMALAB**
In-house analytical result, not an independent third-party test

| LOT / BATCH NUMBER | DATE MADE | DATE OF ANALYSIS | DATE ISSUED |
|---|---|---|---|
| PL-2610-211 | 2026-10-07 | 2026-10-08 | 2026-10-22 |
| **RETEST / EXPIRY** | **UNITS IN LOT / SAMPLED** | **STORAGE** | **CERTIFICATE TYPE** |
| 2028-10-07 | 287 / 2 | — | Lot COA, 2 pages |

| COMPOUND | CAS NUMBER | MOLECULAR FORMULA | THEORETICAL MASS (AVG.) |
|---|---|---|---|
| **Thymogen**<br>Glu-Trp dipeptide (EW); oglufanide | 38101-59-6 | $C_{16}H_{19}N_3O_5$ | 333.34 Da |

SEQUENCE
EW

PURITY · RP-HPLC, AREA %
**99.73%**
Specification ≥ 99.0 % • main peak at 5.24 min


IDENTITY · OBSERVED MASS
**333.33 Da**
Theoretical 333.34 Da • −0.01 Da (30 ppm)

OVERALL CONCLUSION
**PASS**
Meets every specification for the tests performed

| TEST | METHOD | SPECIFICATION | RESULT | CONCLUSION |
|---|---|---|---|---|
| **Identity (mass)** | LC-MS, ESI positive | Theoretical ± 1.0 Da | 333.33 Da | PASS |
| **Purity** | RP-HPLC, UV 214 nm, area % | ≥ 99.0 % | 99.73 % | PASS |
| **Largest impurity** | RP-HPLC, area % | ≤ 1.0% | 0.15% (RT 6.10) | PASS |
| **Assay (label claim)** | Assay vs label claim | 9.5–10.5 mg/mL | 9.90 mg/mL | PASS |
| **Bacterial endotoxin** | LAL, kinetic | ≤ 5.0 EU/mg | 0.190 EU/mg | PASS |
| **Sterility** | USP <71>, 14-day | No growth | No growth | PASS |

## 1 HPLC CHROMATOGRAM

**Data file:** PL-2610-211-HPLC.raw · Column C18, 150 × 4.6 mm · UV 214 nm




Peak plot drawn from the integrated peak table (area %, log scale); it is not the instrument trace, which is retained in PL-2610-211-HPLC.raw. † Nominal reference retention time for the assigned impurity; the main-peak RT is observed in this lot.


**For laboratory research use only.** Not for human or veterinary use.

## 2 HPLC METHOD · PL-2610-211-HPLC.RAW

**INSTRUMENT**
Waters Alliance e2695

**DATA SYSTEM**
Empower 3

**DETECTOR**
UV absorbance

**DETECTION WAVELENGTH**
214 nm

**COLUMN**
C18, 150 × 4.6 mm, 3.5 µm

**MOBILE PHASE A**
0.1 % TFA in water

**MOBILE PHASE B**
0.1 % TFA in MeCN

**GRADIENT**
5–95 % B over 25 min (linear)

**FLOW RATE**
1.0 mL/min

**COLUMN TEMPERATURE**
40 °C

**INJECTION VOLUME**
10 µL

## 3 GRADIENT PROGRAM

| TIME (MIN) | % A | % B |
|---|---|---|
| 0.0 | 95 | 5 |
| 25.0 | 5 | 95 |

## 4 PEAK TABLE

| PEAK | RT (MIN) | AREA | AREA % |
|---|---|---|---|
| 1 | 4.40 † | — | 0.06 |
| **2** | **5.24** | **—** | **99.73** |
| 3 | 6.10 † | — | 0.15 |
| 4 | 6.70 † | — | 0.06 |
| Total | | — | 100.00 |

Peaks: 1 unidentified impurity; 2 main peak (Thymogen); 3 unidentified impurity; 4 unidentified impurity. Area % from the integrated peak table (absolute areas are not part of the lot-release export). † nominal reference RT for the assigned impurity. Purity = main peak area ÷ total integrated area × 100, to two decimals; main peak shaded.

## 5 IDENTITY BY MASS SPECTROMETRY

**INSTRUMENT**
Waters SQD

**IONIZATION / MODE**
ESI, positive

**REFERENCE M/Z (CHARGE)**
334.35 (M+H)+

**DECONVOLUTED MASS**
333.33 Da

**THEORETICAL MASS**
333.34 Da

**DIFFERENCE**
−0.01 Da (30 ppm)

**DATA FILE**
PL-2610-211-MS.d

**ACCEPTANCE**
Within ± 1.0 Da

## 6 ADDITIONAL TESTS

| TEST | PERFORMED BY | DATE | REPORT / FILE |
|---|---|---|---|
| Assay vs label claim | PrymaLab Analytical | 2026-10-08 | — |
| Endotoxin (LAL, kinetic) | PrymaLab Analytical | 2026-10-08 | — |
| Sterility (USP <71>, day-14 read-out) | PrymaLab Analytical | 2026-10-22 | — |

A test with no row here was not performed on this lot. Outside-laboratory reports, when used, are attached and named below.

## 7 REVIEW AND APPROVAL

**D. Reynolds**
Analyst · performed the testing
Date: 2026-10-22

**S. Calloway**
Reviewer · checked against raw data
Date: 2026-10-22

**M. Phelps**
Lab Director · approved for release
Date: 2026-10-22

Raw data retained at: Empower 3, project PL-2610, archive PL-ARC-2026; 7 years. Outside laboratory reports attached: None

This certificate applies only to the lot named above and describes the units sampled from it; it is not a guarantee for any other lot. Results were obtained by PrymaLab LLC's own laboratory (PrymaLab Analytical) and are not an independent third-party test unless an outside report is attached. Reference values (CAS, formula, theoretical mass, sequence) are literature values for the named compound. Authenticity: scan the QR code or enter certificate number PL-26-00323 and key 222GLLFM at prymalab.net/testing/verify; any copy that does not match the record shown there is not a PrymaLab certificate. PrymaLab LLC products are supplied for laboratory research use only and not for human or veterinary use.